

TGR205x

UPDATING FIRMWARE REQUIRES THE USE OF USB FLASH DRIVE

Note that all settings will be returned to their factory defaults, and all un-saved data will be lost. Data and settings stored as files will be unaffected.

Instructions:

- 1) Copy “_TGR205X” folder to the ROOT of a USB Flash-drive
- 2) Connect the USB Flash-drive to the instrument (TGR2053/TGR2051)
- 3) Power-cycle the instrument
- 4) Follow the on-screen instructions